

DECLARATION OF CONFORMITY

according to ISO/IEC Guide 22 and EN 45014

Manufacturer's Name:	Hewlett-Packard Ltd.
Manufacturer's Address:	Queensferry Telecom Operation South Queensferry West Lothian, EH30 9TG Scotland, United Kingdom
Declares that the product	
Product Name :	SDH/PDH/ATM BER Test Set
Model Numbers:	HP 37717 B
Product Options:	This declaration covers all options of the above product.
Conforms to the following Product Specifications:	
Safety:	EN 61010-1 (1993) MEC 1010-1(1990) + A1(1992) CSA C-22.2 No. 1010.1-92 CFR Ch.1 1040.10 Class I EN 60825-1 (1994) For Optical ports only.
EMC:	EN 55011 :1991 Group 1, Class A EN 55082-1:1992
Supplementary Information:	
The product herewith complies with the requirements of the Low Voltage Directive 73/23/EEC and the EMC Directive 89/336/EEC.	
South Queensferry, Scotland	Date 4 Dec 95 W.R. Pearson / Quality Manager
European Contact: Your local Hewlett-Packard Sales and Service Office or Hewlett-Packard GmbH, Department ZQ/Standards Europe Hannemberger Strasse 130, D-7030 Boblingen (FAX: +49-7031-145143)	

Limitation of Warranty	3
Limitation of Remedies and Liability	4
Responsibilities of the Customer	4
CERTIFICATION	5
ASSISTANCE	5
Notice	5
STATEMENT OF COMPLIANCE	5
General	6
Safety Symbols	7
LASER SAFETY	7

1 General Information

Specifications	1-1
Safety Considerations	1-1
Manual Safety Markings	1-2
Instrument Safety Markings	1-2
Laser Safety	1-2
Laser Product Classification	1-3
Laser Warning Symbols	1-4
Swedish and Finnish Labels	1-4
Location of Laser Apertures	1-6
ESD Precautions	1-7
Electromagnetic Compatibility	1-8
Radiated Emissions	1-8
Electrostatic Discharge	1-8
Operators Maintenance	1-9
Cleaning	1-9
Optical Fibre Cleaning	1-9
Instruments Covered By Manual	1-10
Options	1-11

Accessories	1-15
Specifications	1-16
Unframed PDH (Option UKK)	1-16
Structured PDH (Options UKJ, UKL)	1-17
SDH MODULE: Option US1	1-19
OVERHEAD and STRESS TEST SDH MODULE (Options A1T and A3R)	1-21
STM-1/STM-4 OPTICAL INTERFACE: Option UH2	1-26
STM-1/STM-4 OPTICAL INTERFACE: Option URU	1-26
STM-1/STM-4 OPTICAL INTERFACE (1310 nm) with STM-4 Overhead Option UKT	1-27
STM-1/STM-4 (1310/1550 nm) OPTICAL INTERFACE with STM-4 Overhead Option USN	1-28
SDH Binary Interface (Option OYH)	1-30
MEASUREMENTS and ANALYSIS	1-42
UNFRAMED PDH (Option UKK)	1-43
SDH (Option US1)	1-44
SDH (Options A1T, A3R)	1-44
JITTER (Options UHN, A1M, A1N, A1P, A3L, A3V and A3N)	1-45
ATM Measurement (Option UKN)	1-46

2 Installation

Initial Inspection	2-1
Preparation for Use	2-2
Power Requirements	2-2
Power Cord	2-2
Operators Maintenance	2-4
Fuse Replacement	2-4
Line Voltage Selector Switch	2-4
Hewlett-Packard Interface Bus	2-5
HP-IB Address Selection	2-6

Operating Environment	2-7
Storage and Shipment	2-7
Repackaging for Shipment	2-7

3 Performance Tests

Introduction	3-1
Equipment Required	3-1
Extra Modules Required	3-1
Performance Test Record	3-1
Calibration Cycle	3-1
Recall Default Settings	3-2
Self Test - All Options	3-5
Description	3-5
Additional Tests	3-6
PDH Internal Transmitter Clocks (Options UKK, [USB], UKJ, [USA])	3-8
Specifications	3-8
Description	3-8
Equipment Required	3-8
Procedure	3-9
PDH Transmitter Output (UKK, [USB], UKJ, [USA])	3-12
Specifications	3-12
Description	3-12
Procedure	3-13
Multiple PDH Transmitter Outputs (Options UHC, [US6])	3-24
Specifications	3-24
Description	3-24
Procedure	3-25
2.048 Mb/s Output Pulse	3-27
PDH Frame Analysis (Options UKK, [USB])	3-38

Specifications	3-38
Equipment Required	3-38
Description	3-38
Procedure	3-39
Option UKK, [USB] Unstructured PDH	3-39
PDH Receiver Equalization (Options UKK, [USB], UKJ, [USA], UKL, [USC], UKN, [USE])	3-46
Specifications	3-46
Description	3-46
Procedure	3-47
PDH Receiver Monitor Levels (Options UKK, [USB])	3-50
Specification	3-50
Description	3-50
Procedure	3-50
SPDH Receiver Monitor Levels (Options UKJ, [USA], UKL, [USC], UKN, [USE])	3-53
Specifications	3-53
Description	3-53
Procedure	3-54
External 2Mb/s Mux/Demux (Options UKJ, [USA], UKN, [USA])	3-58
Specifications	3-58
Description	3-58
Equipment Required	3-59
Procedure	3-59
External 2Mb/s Demux (Options UKL, [USC])	3-65
Specifications	3-65
Description	3-65
Procedure	3-66
PDH Receiver Monitor Levels (Special Option 808)	3-71
Specification	3-71
Description	3-71
Procedure	3-71

PDH Error Output (Options UKK, [USB])	3-74
Specifications	3-74
Description	3-74
Procedure	3-74
PDH Frequency Measurement and Looped Clock (Options UKK, [USB], UKJ, [USA], UKL, [USC], UKN, [USE])	3-76
Specifications	3-76
Description	3-76
Procedure	3-76
Internal SDH Transmitter Clock (Options US1, [US5], A1T, [A1U], A3R [A3S])	3-80
Specification	3-80
Description	3-80
Procedure	3-80
SDH Frequency Offsets (Options US1, [US5], A1T, [A1U], A3R [A3S])	3-82
Specifications	3-82
Description	3-82
Procedure	3-82
External MTS Clock (Options US1, [US5], A1T, [A1U], A3R [A3S])	3-85
Specifications	3-85
Description	3-85
Procedure	3-85
External MTS Clock (Options US1, [US5], A1T, [A1U], A3R [A3S])	3-88
Specifications	3-88
Description	3-88
Procedure	3-88
STM-1 Transmitter Output Waveshape (Options US1, [US5], A1T, [A1U], A3R [A3S])	3-91
Specifications	3-91
Description	3-91

Procedure	3-91
STM-1 Receiver Monitor Input (Options US1, [US5])	3-97
Specification	3-97
Description	3-97
Procedure	3-97
STM-1 Receiver Monitor Input (Options A1T, [A1U, A3R [A3S]])	3-99
Specification	3-99
Description	3-99
Procedure	3-99
STM-1 Receiver Monitor Input (Special Option 808)	3-101
Specification	3-101
Description	3-101
Procedure	3-101
STM-1 Receiver Input Equalization (Options US1, [US5], A1T, [A1U], A3R [A3S])	3-103
Specification	3-103
Description	3-103
Procedure	3-103
STM-1 Optical Interface (Option UH1)	3-104
Specification (Transmitter)	3-104
Specification (Receiver)	3-104
Description	3-104
Procedure	3-105
STM-1/STM-4 Optical Interface (Option UH2)	3-107
Specification (Transmitter)	3-107
Specification (Receiver)	3-107
Description	3-107
Procedure	3-108
STM-1/STM-4, 1550nm Optical Interface (Option URU)	3-110
Specification (Transmitter)	3-110
Specification (Receiver)	3-110
Description	3-110

Procedure	3-111
STM-1/STM-4 Optical Interface (Options UKT, USN)	3-116
Specification (Transmitter)	3-116
Specification (Receiver)	3-116
Description	3-116
Procedure Option UKT and USN	3-118
Procedure Option USN only	3-120
Transmitted Jitter Amplitude Accuracy (Option UHK)	3-124
Specifications	3-124
Description	3-125
Procedure	3-126
Transmitted Jitter/Wander Amplitude Accuracy (Option A3K [A3Q])	3-131
Specifications	3-131
Wander Timing Reference Input	3-132
External Jitter Modulation Input	3-133
Description	3-133
Pre-Adjustment Setup	3-134
Procedure	3-135
Received Jitter Accuracy (Options UHN, [US9], A1M [A1Q], A1N [A1R], A1P [A1S], A3L [A3M], A3V [A3W], A3N [A3P])	3-148
Specifications	3-148
Description	3-148
Received Jitter Accuracy Procedure	3-149
Wander/Slips Measurement (Options UHN, [US9], A1M [A1Q], A1N [A1R], A1P [A1S], A3L [A3M], A3V [A3W], A3N [A3P])	3-152
Specifications	3-152
Description	3-153
Procedure	3-153
Jitter Receiver STM-1E Input Equalization (Options A1M [A1Q], A1N [A1R], A1P [A1S], A3L [A3M], A3V [A3W], A3N [A3P])	3-158
Specification	3-158

Description	3-158
Procedure	3-158
Jitter Receiver STM-1E Monitor Input (Options A1M [A1Q], A1N [A1R], A1P [A1S], A3L [A3M], A3V [A3W], A3N [A3P])	3-159
Specification	3-159
Description	3-159
Procedure	3-159
Jitter Receiver STM-1 Optical Interface (Options A1N, [A1R], A3V [A3W])	3-161
Specification	3-161
Description	3-161
Procedure	3-162
Jitter Receiver STM-1/STM-4 Optical Interface (Options A1P, [A1S], A3N [A3P])	3-164
Specification	3-164
Description	3-164
Procedure	3-165
PDH Binary Interfaces (Option UH3, [US7])	3-167
Specifications	3-167
Description	3-168
Procedure	3-168
Performance Test Record	3-175
Retrofit Procedure	B-4
Testing	B-5
Jitter Transmitter Module Calibration	B-6
Description	B-6
Equipment	B-6
Pre-Adjustment Setup	B-6
Jitter Transmitter Calibration Procedure	B-7

General Information

This section contains general information concerning the HP 37717B PDH/SDH test set.

SPECIFICATIONS

SAFETY CONSIDERATIONS

ELECTROMAGNETIC COMPATABILITY

OPERATORS MAINTENANCE

INSTRUMENTS COVERED BY MANUAL

OPTIONS

ACCESSORIES

Specifications

The instrument specifications are listed on Pages 1-16 to 1-48. These specifications are the performance standard or limits, against which the instrument is tested.

The recommended test equipment required to carry out the performance tests is listed in Table 1-1.

Safety Considerations

The HP 37717B PDH/SDH test set is a Safety Class 1 (IEC) product (provided with a protective earthing ground incorporated in the power cord).

This instrument has been designed and tested in accordance with IEC Publication 1010-1(1990) with Amendment 1 (1992), Safety Requirements for Electronic Measuring Apparatus, and has been supplied in a safe condition. The instruction documentation contains information and warnings which must be followed by the user to ensure safe operation and to maintain the instrument in a safe condition.